

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-21937-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: **21.12.2023**

Date of issue: 21.12.2023

Holder of accreditation certificate:

LABORATORIO METROLOGIA RADIACIONES IONIZANTES CCHEN
Avda. Nueva Bilbao Nº 12.501, 760071 LAS CONDES, REGIÓN METROPOLITANA
SANTIAGO DE CHILE, CHILE

with the location

LABORATORIO METROLOGIA RADIACIONES IONIZANTES
Avda. Nueva Bilbao Nº 12.501, 760071 LAS CONDES, REGIÓN METROPOLITANA
SANTIAGO DE CHILE, CHILE

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the general with the principles of DIN EN ISO 9001.

Calibrations in the fields:

High frequency and radiation quantities

- Ionising radiation and radioactivity**
 - **Dosimetry**

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded measurement uncertainty	Remarks
Air kerma	1 Gy to 5 Gy	Cobalt 60	2 %	
Absorbed dose to water				

Abbreviations used:

DIN Deutsches Institut für Normung e.V. – German institute for standardization